


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVPP-1-PS-P-PN-M-CA-ED-EN*

S.N.: *4382*

Date: *Aug 4 88*



(A) *A out*
40 dB ATTEN
10 VOLTS/DIV
500 kHz PRF



(B) *B out*
40 dB ATTEN
10 VOLTS/DIV
500 kHz PRF

- a) Output signal amplitude:
0 TO ±10 VOLTS
- b) Pulse width:
0.2 TO 100 NS
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 10 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz

[Signature]