


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVL-2-PS-P  
S.N.: 4384  
Date: JUNE 9 88



- a) Output signal amplitude:  
0 TO +350 VOLTS
- b) Pulse width:  
5 TO 100 NSEC.
- c) Rise time:  
≤ 2 NSEC
- d) Fall time:  
≤ 3 NSEC
- e) PRF:  
0 TO 5 KHz
- f) Jitter, stability:  
OK
- g) Prime power:  
120/240 V  
50-60 Hz.

70 dB ATTEN

320 VOLTS/DIV

2 KHz PRF