

PULSE GENERATOR
PERFORMANCE CHECK

Model: AVL-2-PS-P.

S.N.: 4387

Date: JUNE 9, 1988



- a) Output signal amplitude:
0 TO +350 VOLTS
- b) Pulse width:
5 TO 100 NSEC
- c) Rise time:
 ≤ 2 NSEC
- d) Fall time:
 ≤ 3 NSEC
- e) PRF:
0 TO 5 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz

70 db ATTEN

320 VOLTS/DIV

2 KHz PRF