


PULSE GENERATOR

PERFORMANCE CHECK


Model: *PRR-B3-C-PN-W-05*

S.N.: *4394*

Date: *JULY 20 1988*



Ⓐ P OUT  
60 dB ATTEN  
100 VOLTS/DIV  
5.0 KHZ



Ⓑ N OUT.  
 $R_L = 50 \Omega$   
5.0 KHZ

- a) Output signal amplitude:  
*0 TO ± 100 VOLTS*
- b) Pulse width:  
*0.1 μSEC TO 5.0 μSEC*
- c) Rise time:  
*≤ 5 NSEC*
- d) Fall time:  
*≤ 5 NSEC*
- e) PRF:  
*0 TO 5 KHZ*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240V*  
*50-60 HZ*

*[Signature]*