


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVO-1-PS-P*

S.N.: *4395*

Date: *JUN 22 88*



- a) Output signal amplitude:
+25 TO -100 AMPS
- b) Pulse width:
≤ 25 NPER
- c) Rise time:
NA
- d) Fall time:
2A
- e) PRF:
0 40 1.0 KHz
- f) Jitter, stability:
OK
- g) Prime power:

*≈ 60 AMP/DIV
500 KHz PRF*

*120/240 V
50-60 Hz*