


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AK-P*
S.N.: *4401*
Date: *JULY 20 1988*



- a) Output signal amplitude:
≥ 115 VOLTS
- b) Pulse width:
7 TO 100 NS
- c) Rise time:
2100 PSEC
- d) Fall time:
≤ 2 NS
- e) PRF:
MIN PW: 0 TO 1.0 MITE
MAX PW: 0 TO 20 KITE
- f) Jitter, stability:
OK
- g) Prime power:

40 db ATTEN
10 VOLTS DIV
1.0 MITE

115 ± 1.0 VOLT
150 mA (MAX)