


PULSE GENERATOR
PERFORMANCE CHECK

Model: MK-P.

S.N.: 4402

Date: JULY 20 1988



- a) Output signal amplitude:
 ≈ 15 VOLTS.
- b) Pulse width:
 4 TO 100 nSEC
- c) Rise time:
 ≤ 100 pSEC
- d) Fall time:
 ≤ 2 nSEC
- e) PRF:
 MIN PW: 0 TO 1.0 MHz
 MAX PW: 0 TO 20 kHz
- f) Jitter, stability:
 OK
- g) Prime power:

40dB ATTEN
10 VOLTS/DIV
1.0 MHz

15 \pm 1.0 VOLTS
150 mA (MAX)