


PULSE GENERATOR
PERFORMANCE CHECK

Model: *APL-2-C-N-GEN-05-PM-M1*

S.N.: *4404*

Date: *Aug 5 88*



- a) Output signal amplitude:
0 TO - 350. V
- b) Pulse width:
5 TO 100 NS
- c) Rise time:
≤ 2 NS
- d) Fall time:
≤ 3 NS
- e) PRF:
0 TO 5 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V

TO 26 ATTEN
≈ 320 VOLTS/DIV
2 KHz PRF

50-60 Hz