


PULSE GENERATOR
PERFORMANCE CHECK


Model: *APPP-1-PS-P-PN-N-EX-ES-EO*

S.N.: *4413*

Date: *JUNE 23 88*



A out



B. out

- a) Output signal amplitude:
0 TO ± 10 VOLTS
- b) Pulse width:
0.2 TO 100 NSEC
(TWO RANGES)
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120 / 240 V
50 ~ 60 MHz.

40 db ATTEN

10 VOLTS/DIV

1.0 MHz PRF