


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVPP-1-PS-P-PN-M-EA-EW-ED*
S.N.: *4415*
Date: *JUNE 23 88*



A out



B out

40 dB ATTEN

10 VOLTS/DIV

1.0 MHz PRF

- a) Output signal amplitude:
0 TO ± 10 VOLTS
- b) Pulse width:
0.2 TO 100 NS
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
- g) Prime power:
OK

120/240 V

50-60 MHz