


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVB1-C-PM4


S.N.: 4416

Date: JUNE 21 80



- a) Output signal amplitude:  $\approx 40$  VOLTS
- b) Pulse width: 1 NSR (1 BIT PERIOD)
- c) Rise time: NA
- d) Fall time: NA
- e) PRF: 0 TO 100 KHz
- f) Jitter, stability: OK
- g) Prime power: 120/240 V  
50-60 Hz

① 40 dB ATTN  
10 VOLTS/DIV  
100 KHz PRF



② AS ① BUT  
500 PSER/DIV

*[Signature]*