

PULSE GENERATOR
PERFORMANCE CHECK

Model: AVH-P-IMA
S.N.: 4423
Date: July 19 88

40 db atten
10 volts/div
5.00 kHz

- a) Output signal amplitude:
0 to +30 volts
- b) Pulse width:
≤ 1 nsec
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 to 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 ± 1.0 volt
150 mA (max)
Mf.