


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PAH-N-IM*
S.N.: *4428*
Date: *JULY 19 88*



40 dB ATTEN
10 VOLT/DIV
500 MHz

- a) Output signal amplitude:
0 TO -30 VOLTS
 - b) Pulse width:
≤ 1 NS
 - c) Rise time:
NA
 - d) Fall time:
NA
 - e) PRF:
0 TO 1.0 MHz
 - f) Jitter, stability:
OK
 - g) Prime power:
15 ± 1.0 VOLTS
150 mA (MAX)
- [Signature]*