


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AMR-3PW-C-P*
S.N.: *4430*
Date: *JULY 6 1988*



Ⓐ $R_c = 50 \Omega$
 $PRF = 50 \text{ Hz}$



Ⓑ $PRF = 1.0 \text{ kHz}$

- a) Output signal amplitude:
0 TO + 200 VOLTS
- b) Pulse width:
0.1 TO 100 USEC
- c) Rise time:
≤ 10 NSEC
- d) Fall time:
≤ 10 NSEC
- e) PRF:
MIN PW: 0 TO 50 Hz
PW ≤ 5 USEC: 0 TO 1 kHz
- f) Jitter, stability:
OK
- g) Prime power:
120 / 240 V
50 - 60 Hz