


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AD-2W-C-N-PN-MA-05*
S.N.: *4431*
Date: *AUG 8 1988*



- a) Output signal amplitude:
0 TO ± 10 AMPS
- b) Pulse width:
2 TO 50 NSEC
- c) Rise time:
≤ 2 NSEC
- d) Fall time:
≤ 1 NSEC
- e) PRF:
0 TO 20 KHZ
- f) Jitter, stability:

~ 7 AMPS/DIU

20 KHZ PRF

- g) Prime power:
OK
120 / 240V
50 ~ 60 KHZ