


PULSE GENERATOR
PERFORMANCE CHECK

Model: *ANMR-1-P-CA-ED-ED-TRF-ECL-MMI-JM*

S.N.: *4433*

Date: *AUG 22 88*



- a) Output signal amplitude:
0 TO +5 VOLTS
- b) Pulse width:
5 TO 50 nSEC
- c) Rise time:
20% MAX DUTY CYCLE
0.1 OR 1.0 nSEC
- d) Fall time:
0.1 OR 1.0 nSEC
- e) PRF:
0 TO 10 MHz
(DUTY CYCLE MAX 20%)
- f) Jitter, stability:
OK

40 dB ATTEN
5 VOLTS/DIV
1.0 nTR

TR, TF LOW

- g) Prime power:
+24 V ± 1.0 V
300 mA (MAX)
- h) OFFSET: *0 TO ±5 V*
TO 50 Ω

AA