

PULSE GENERATOR
PERFORMANCE CHECK

Model:


PR-1-P-033

S.N.:

4440

Date:

JULY 6 1988



$$R_L = 50 \Omega$$

$$PRF = 20 \text{ kHz}$$

- a) Output signal amplitude:
0 to +200 Volts to
- b) Pulse width: 30 ns
20 nsec to 100 nsec
- c) Rise time:
 $\leq 10 \text{ nsec}$
- d) Fall time:
 $\leq 10 \text{ nsec}$
- e) PRF: 0 to 20 kHz.
- f) Jitter, stability:
OK
- g) Prime power:

+24 Volts
500 mA (max).