

PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-5B-C-PN-ES1*


S.N.: *4465*

Date: *AUG 15 1988*



- a) Output signal amplitude:
0 TO ± 40 AMPS
- b) Pulse width:
0.2 TO 2.0 USEC
- c) Rise time:
≤ 10 NSEC
- d) Fall time:
≤ 10 NSEC
- e) PRF:
0 TO 1.0 KHZ
- f) Jitter, stability:
OK

Ⓐ *P_{out}*
R_L = 1.0 Ω
PRF = 1.0 KHZ



- g) Prime power:
120 / 240 V
50 - 60 MHz
- [Signature]*

Ⓑ *N_{out}*
R_L = 1.0 Ω
PRF = 1.0 KHZ