


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *APR-B2-C-CA-PN*

S.N.: *4471*

Date: *April 25 1988*



- a) Output signal amplitude:  
*0 TO ±50 VOLTS*
- b) Pulse width:  
*20 NSER TO 200 NSER*
- c) Rise time:  
*≤ 3 NSER*
- d) Fall time:  
*≤ 3 NSER*
- e) PRF:  
*0 TO 50 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*50 dB ATTEN*

*32 VOLTS/DIV*

*50 KHz*

*120/240V*

*50-60 Hz*