


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVL-2-C-T-PN-OS-M.*

S.N.: *4477*

Date: *April 30 1988*



- d) Output signal amplitude:
0 to ± 350 V.
- e) Pulse width:
5 to 100 nsec
- f) Rise time:
≤ 2 to 10 nsec
- g) Fall time:
3 to 10 nsec
- h) PRF:
0 to 5 kHz
- i) Jitter, stability:
OK
- j) Prime power:

70 db return

2 320 VOLTS/DIV

2 kHz

120/240 V

50-60 Hz

[Signature]