


PULSE GENERATOR
PERFORMANCE CHECK


Model: *APP-1-C-GA-EN-N-N*

S.N.: *4480*

Date: *AUG 29 88*



*(A) out. 40 db ATTEN
1.0 MHz*



*(B) out
40 db ATTEN
1.0 MHz*

- a) Output signal amplitude:
0 TO = 10 V. 0.483
- b) Pulse width:
0.2 TO 100 NSER.
- c) Rise time:
≤ 100 PSEZ
- d) Fall time:
≤ 135 PSEZ
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
*120/240 V
50-60 Hz*

[Signature]
7