


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *PK-MV-N-JM*  
S.N.: *4502*  
Date: *SEPT 8 88*



- a) Output signal amplitude:  
*0 TO -15 VOLTS*
- b) Pulse width:  
*1 TO 50  $\mu$ SEC*
- c) Rise time:  
 *$\leq$  100 PSEC*
- d) Fall time:  
 *$\leq$  1  $\mu$ SEC*
- e) PRF:  
*0 TO 50 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*+15  $\pm$  1.0 VOLTS*

*40 db ATEN*  
*10 VOLTS DIV*  
*20 KHz PRF*

*250 mA (max)*  
*AS*