


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVK-MV-N-DM*

S.N.: *4503*

Date: *SEPT 8 '88*



- a) Output signal amplitude:
0 TO -15 VOLTS
- b) Pulse width:
1 TO 50 NS
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 1 NS
- e) PRF:
0 TO 50 KHz
- f) Jitter, stability:
OK
- g) Prime power:

40 db ATTEN
10 VOLTS/DIV
20 KHz PRF

+15 ± 1.0 VOLTS
250 mA (MAX)