


PULSE GENERATOR
PERFORMANCE CHECK

Model: *MP-10-14V3-C-P-TM*

S.N.: *4515*

Date: *SEPT 14 '88*



- a) Output signal amplitude: *0 TO +40 VOLTS*
- b) Pulse width: *0.4 TO 2.0 NS*
- c) Rise time: *≤ 150 PS*
- d) Fall time: *≤ 250 PS*
- e) PRF: *0 TO 1.0 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *170/240V*

50 dB ATTEN

≈ 32 VOLTS / DIV

100 kHz PRF

170/240V

50-60 Hz