


PULSE GENERATOR  
PERFORMANCE CHECK


Model: *AVM-1C-D-M-OT-T*

S.N.: *4556*

Date: *SEPT 19 88*



Ⓐ 40 dB ATTEN  
5 VOLTS/DIV  
PRF = 25 MHz



Ⓑ AS Ⓐ BUT  
1 NS/DIV

- a) Output signal amplitude:  
*0 TO +5 VOLTS*
- b) Pulse width:  
*0.2 TO 6.0 NS*
- c) Rise time:  
*≤ 100 PSEC*
- d) Fall time:  
*≤ 135 PSEC*
- e) PRF:  
*0 TO 25 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120 / 240 V*  
*50-60 Hz.*

*[Signature]*

- h) OFFSET:  
*0 TO ±5 VOLTS*  
*TO 50 ~*