


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AV6-3-C-P-PW-K52*



S.N.: *4558*

Date: *FEB 8 '89*



- a) Output signal amplitude:
0 TO ± 650 VOLTS
- b) Pulse width:
≤ 5 nSEC
- c) Rise time: (20% RISE TIME)
≤ 2 nSEC
- d) Fall time:
≤ 2 nSEC
- e) PRF:
0 TO 20 KHZ
- f) Jitter, stability:
OK
- g) Prime power:
120 / 240 V, 50-60 HZ

(A)
P_{OUT} 60 dB ATTEN
100 VOLTS/DIV
PRF = 10 KHZ



(B) AS *(A)* BUT
2 nSEC/DIV

(C) N_{OUT} (A WITH
AUX-1).