


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *PL-2-C-N*

S.N.: *4588*

Date: *OCT 27 88*



- a) Output signal amplitude:  
*0 TO -350 V*
- b) Pulse width:  
*5 TO 100 nSEC*
- c) Rise time:  
*≤ 2 nSEC*
- d) Fall time:  
*≤ 3 nSEC*
- e) PRF:  
*0 TO 5 kHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120 / 240 V*  
*50-60 Hz*

*70 db ATTEN*  
*≈ 320 VOLTS/DIV*  
*PRF = 2.0 kHz*  
*R<sub>L</sub> = 50 Ω*

*[Signature]*