

PULSE GENERATOR  
PERFORMANCE CHECK

Model:


AVMN-1-e-p

S.N.:


4609

Date:

DEC 1 88



(A) 40 db ATTEN  
5 VOLT/DIV  
~ 25 MHz



(B) 40 db ATTEN  
5 VOLT/DIV  
~ 50 MHz

a) Output signal amplitude:

0 TO +5 VOLTS

b) Pulse width:

0.3 TO 5.0 NS

c) Rise time:

~ 100 PSEC

d) Fall time:

~ 200 PSEC

e) PRF:

0 TO 50 MHz

f) Jitter, stability:

OK

g) Prime power:

120/240 V

50-60 Hz