


PULSE GENERATOR
PERFORMANCE CHECK

Model: *M14-MPT-P-TM*

S.N.: *4632*

Date: *DEC 15 1988*



50 dB ATTEN
≈ 32 VOLTS/DIV
500 KHZ

- a) Output signal amplitude: *≥ +30 VOLTS*
- b) Pulse width: *≤ 1 nSEC*
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 TO 1 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *+15 ± 1.0 VOLTS*
150 mA (max)