


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVH-M2-P*  
S.N.: *4637*  
Date: *DEC 6 '88*



*50 db ATTEN*  
*≈ 32 VOLTS/DIV*  
*1.0 MHz*  
*PRT*

- a) Output signal amplitude:  
*≥ +60V*
- b) Pulse width:  
*≤ 2.0 NSER*
- c) Rise time:  
*NA*
- d) Fall time:  
*NA*
- e) PRF:  
*0 TO 1 MHz*
- f) Jitter, stability:
- g) Prime power:  
*OK*  
*+135 ± 5 VOLTS*  
*5 mA (MAX)*

*[Signature]*