


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AMR-1-pw-ps-pn*

S.N.: *4657*

Date: *NOV 29 '83*



*ⓐ P_{out}
R_L = 50Ω*



ⓑ N_{out}

- a) Output signal amplitude:
0 TO ± 200 V.
- b) Pulse width:
0.1 TO 5.0 μSEC
- c) Rise time:
≤ 10 nSEC
- d) Fall time:
≤ 10 nSEC
- e) PRF:
0 TO 1 KHz
- f) Jitter, stability:
OK
- g) Prime power:

*120/240V
50-60Hz*