


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AV14-141-MA2*

S.N.: *4664*

Date: *JAN 12 '89*



*60 db ATEN*

*5.0 V/DIV*

*20 kHz PRF*

a) Output signal amplitude:

*> +100 V RMS*

b) Pulse width:

*≤ 650 psec*

c) Rise time:

*NA*

d) Fall time:

*NA*

e) PRF:

*0 TO 20 kHz*

f) Jitter, stability:

*OK*

g) Prime power:

*+15 ± 1.0 V RMS*

*150 mA (MAX)*

*[Signature]*