


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PMH-S-MVI-P*

S.N.: *4667*

Date: *JAN 12 89*



40 db ATTN

10 V/DIV
500 KHz (PRF)

- a) Output signal amplitude:
0 TO +40 VOLTS
- b) Pulse width:
≤ 350 psec (FWHM)
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 ± 1.0 VOLTS
150 mA (MAX)

[Signature]