


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVMA-1-P-EN*

S.N.: *4696*

Date: *DEC 14 80*



40 dB ATTEN

5 VOLT/DIV

≈ 7.0 MHz PRF

- a) Output signal amplitude:
0 TO +5 VOLT
- b) Pulse width:
0.2 TO 6.0 NSEC
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:
+24 ± 1.0 VOLT
400 mA (MAX)

M