


PULSE GENERATOR

PERFORMANCE CHECK


Model: *PAK-32-C-W-TR-CA-EW*
-P-PN-ATT2

S.N.: *4701*


Date: *FEB 9 89*



*(A) 50 db ATTEN. 32 VOLTS/DIV
RISE TIME OF P*




(B) AS (A) BUT N



*(C) AS (B) BUT
50 NS/DIV*

- a) Output signal amplitude:
0 TO ±50 VOLTS
- b) Pulse width:
0.1 TO 5.0 NS
- c) Rise time:
≤ 0.5 NS
- d) Fall time:
≤ 5 NS
- e) PRF:
0 TO 5 KHz
- f) Jitter, stability:
OK
- g) Prime power:
*120 / 240 V
50-60 Hz*



*(D) PONT R_{in} = 50Ω
10 VOLTS/DIV
1.0 NS/DIV*