


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVD-AV-14V2-C-PN-05-IM*

S.N.: *4707*

Date: *FEB 7 89*



*40 db ATTEN*  
*10 VOLTS/DIV*  
*100 KHZ PRF*

- a) Output signal amplitude:  
*0 TO ±20 VOLTS*
- b) Pulse width:  
*0.4 TO 2.0 NS*
- c) Rise time:  
*≤ 150 PSEC*
- d) Fall time:  
*≤ 250 PSEC*
- e) PRF:  
*0 TO 1.0 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240 V*  
*50-60 Hz*