



PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVR-3-PN-0A-KS2*
S.N.: *4723*
Date: *JAN 11 '89*



Ⓐ $R_L = 50 \Omega$ PRF = 200 Hz



Ⓑ $R_L = 50 \Omega$ PRF = 50 Hz

- a) Output signal amplitude:
0 TO ±200 VOLTS
- b) Pulse width:
0.2 TO 250 μS
- c) Rise time:
≤ 10 NS
- d) Fall time:
≤ 10 NS
- e) PRF:
0 TO 10 Hz
- f) Jitter, stability:
OK
- g) Prime power:
+24V, 1.0 AMP MAX.