


PULSE GENERATOR  
PERFORMANCE CHECK


Model: *AVRL-1TT9B-05*

S.N.: *4750*

Date: *MAR 30 89*



*A out*



*B out.*

- a) Output signal amplitude:  
*A) -200V*  
*B) -200V*
- b) Pulse width:  
*A) 70 TO 100 NS*  
*B) 100 NS TO 1.0 US*
- c) Rise time:  
*≤ 5 NS*
- d) Fall time:  
*≤ 5 NS*
- e) PRF:  
*0 TO 1 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120 / 240 V*  
*50 - 60 Hz*