


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AKR-3-PW-PS-PN*

S.N.: *4755*

Date: *MAR 31 89*



*Power $R_L = 50 \Omega$
PRF = 50 Hz*



*Power $R_L = 50 \Omega$
PRF = 1.0 KHz*

- a) Output signal amplitude:
0 to ± 200 V.
- b) Pulse width:
0.1 to 100 μ s
- c) Rise time:
 ≤ 10 ns
- d) Fall time:
 ≤ 10 ns
- e) PRF:
*0 - 50 Hz (MITX PW,
0 - 1 kHz (PW $\leq 1.0 \mu$ s)*
- f) Jitter, stability:
OK
- g) Prime power:
*120/240 V
50-60 Hz.*