


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVO-7B-C-P-5NLY*


S.N.: *4759*

Date: *APRIL 7 89*



- a) Output signal amplitude:  
*0 TO +100 AMPS*
- b) Pulse width:  
*TO  $\approx 1.0 \mu s$*
- c) Rise time:  
*TO 200  $\mu s$*
- d) Fall time:  
 *$\leq 1.0 \mu s$*
- e) PRF:  
 *$\leq 2.0 \mu s$*
- f) Jitter, stability:  
*0 TO 3 Hz AT MAX DUTY CYCLE*

Ⓐ *R<sub>L</sub>  $\approx 0.8 \Omega$   
20 VOLTS/DIV  
10  $\mu s$ /DIV*



- g) Prime power:  
*OK*  
*120/240 V*  
*50-60 Hz*

Ⓑ *R<sub>L</sub>  $\approx 0.8 \Omega$   
20 VOLTS/DIV  
50  $\mu s$ /DIV*

*[Signature]*