


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVB-5B-CP*

S.N.: *4761*

Date: *APRIL 26 '89*



$$R_c = 0.7 \Omega$$

$$\therefore \approx 1.43 \text{ Amp/DIV}$$

$$PRF = 10 \text{ KHz}$$

a) Output signal amplitude:

0 TO +48 AMPS

b) Pulse width:

25 TO 500 NS

c) Rise time:

≤ 10 NS

d) Fall time:

≤ 10 NS

e) PRF:

0 TO 1.0 KHz

f) Jitter, stability:

OK

g) Prime power:

$$120/240 \text{ V}$$

$$50-60 \text{ Hz}$$