


PULSE GENERATOR
PERFORMANCE CHECK


Model: *MVI-U-HV2-C-P*

S.N.: *4765*

Date: *MARCH 22 89*



*Ⓐ 60 db atten
100 volts/DIV*



*Ⓑ AS Ⓐ BUT
1 nS/DIV
(RISE TIME)*

- a) Output signal amplitude:
0 TO 700 VOLTS
- b) Pulse width:
5 TO 100 nSEC
- c) Rise time:
≤ 0.5 nSEC
- d) Fall time:
≤ 7 nSEC
- e) PRF:
0 TO 5 KHz
- f) Jitter, stability:
OK
- g) Prime power:
170/240 V
50 to 170

[Signature]