


PULSE GENERATOR
PERFORMANCE CHECK


Model: *ANM-3-C-P-PN*

S.N.: *4760*

Date: *APRIL 25 '09*



Ⓐ *40 db ATTEN*
10 VOLTS/DIV
≈ 1.0 nsec



Ⓑ *40 db ATTEN*
10 VOLTS/DIV
≈ 25 nsec.

- a) Output signal amplitude:
0 TO ±15 VOLTS
- b) Pulse width:
2 TO 15 nsec
- c) Rise time:
≤ 150 psec
- d) Fall time:
≤ 600 psec.
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240V
50-60 Hz.

[Signature]