


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *APR-B3-PS-P-CA-EW*  
S.N.: *4773*  
Date: *APRIL 10 89*



*$R_L = 50 \Omega$*   
*PRF = 2 kHz*

- a) Output signal amplitude:  
*0 TO +100 VOLTS*
  - b) Pulse width: *TO 50 ns*  
*0.1 TO 5.0  $\mu$ SER*
  - c) Rise time:  
 *$\leq 5$  NSER*
  - d) Fall time:  
 *$\leq 5$  NSER*
  - e) PRF:  
*0 TO 5 kHz*
  - f) Jitter, stability:  
*OK*
  - g) Prime power:  
*120/240 V*  
*50-60 Hz*
- [Signature]*