


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *ANL-2-PS-P*

S.N.: *4785*

Date: *Jan 6 89*



- a) Output signal amplitude:  
*0 TO +350. V*
- b) Pulse width:  
*5 TO 100 nSec*
- c) Rise time:  
*5.2 nSec*
- d) Fall time:  
*5.3 nSec*
- e) PRF:  
*0 TO 5 kHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240V*  
*50-60 Hz*

*70 db 1572V*  
*320 VOLTS/DIV*  
*2.0 kHz*