


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVL-2-P5-P*  
S.N.: *4787*  
Date: *MAR 10 '89*



- a) Output signal amplitude:  
*0 TO +350 V*
- b) Pulse width:  
*5 TO 100 NS*
- c) Rise time:  
*≤ 2 NS*
- d) Fall time:  
*≤ 3 NS*
- e) PRF:  
*0 TO 5 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240 V*

*70 db ATTEN*  
*≈ 320 V<sub>0</sub> / DIV*  
*PRF ≈ 2 KHz*

*30-60 Hz*

*[Signature]*