


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVL-2-PS-P .

S.N.: 4789

Date: MAR 10/89



- a) Output signal amplitude:
0 TO +350V
- b) Pulse width:
5 TO 100 NSEC
- c) Rise time:
 ≤ 2 NSEC
- d) Fall time:
 ≤ 3 NSEC
- e) PRF:
0 TO 5 KHz
- f) Jitter, stability:
OK
- g) Prime power:

70 db ATTEN
 ≈ 320 VOLTS/DIV
PRF ≈ 2 KHz

120/240V
50-60 Hz