



PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVR-4B-PW-C-PN*  
 S.N.: *4813*  
 Date: *APRIL 13 89*



- a) Output signal amplitude:  
*0 TO ±400 VOLTS*
- b) Pulse width:  
*0.1 TO 100 USEC*
- c) Rise time:  
*≤ 20 NSEC*
- d) Fall time:  
*≤ 20 NSEC*
- e) PRF:  
*MAX DW: 0 TO 50 Hz*  
*PW ≤ 10 μsec: 0 TO 1 KHz*
- f) Jitter, stability:

*P<sub>out</sub> R<sub>L</sub> = 50 Ω*  
*PRF = 1.0 KHz*



- g) Prime power:  
*OK*  
*120 / 280 V,*  
*50-60 Hz*

*N<sub>out</sub> R<sub>L</sub> = 50 Ω*  
*PRF = 50 Hz*