



PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVM-3-N*  
S.N.: *4821*  
Date: *MAY 2 89*



(A) *40 db ATTEN*  
*10 VOLTS/DIV*  
*PRT ≈ 7.0 MHz*



(B) *10 VOLTS/DIV*  
*PRT ≈ 25 MHz*

- a) Output signal amplitude:  
*0 TO -15V*
- b) Pulse width:  
*2 TO 15 NSEC*
- c) Rise time:  
*≤ 150 PSEC*
- d) Fall time:  
*≤ 600 PSEC*
- e) PRF:  
*0 TO 25 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*+28V*  
*400 mA (MAX)*