


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-3A-CP*

S.N.: *4890*

Date: *JULY 18 1989*



a) Output signal amplitude:
+25 TO +75 AmB

b) Pulse width:
≤ 2.5 NS

c) Rise time:
NA

d) Fall time:
NA

e) PRF:
0 TO 20 KHz

f) Jitter, stability:
OK

g) Prime power:
120/240V
50-60 Hz

R_L ≈ 5.0 Ω
60 dB ATTEN

≈ 100 VOLT/DIV
≈ 20 AMP/DIV

[Signature]